

Product XC air temperature sensor
Serial number 8H1451
Date of calibration 29 April 2020
Customer name
Customer address
Customer reference



Calibration certificate

Specification Air temperature sensor accuracy ± 0.20 °C

Measured values and uncertainties of calibration

Time (mins)	Air temperature (°C)		
	Reference	Measured	Error
0	21.90	21.88	-0.02
10	21.90	21.89	-0.01
20	21.91	21.89	-0.02
30	21.91	21.91	0.00
40	21.92	21.91	-0.01
50	21.93	21.92	-0.01
60	21.94	21.92	-0.02
Mean error			-0.02
Uncertainty (k=2)			0.10

Reference standards	Ref. no.	Certificate no.	Date
XC air temperature sensor	7N5247	2021033091-LL01-3	30 March 21
Test Procedure	PT-50304-02		11 April 19

Laser measurement system accuracy: When a Renishaw XL-80 laser is used with a Renishaw XC-80 compensator and a Renishaw air temperature sensor (*all within specification*) the laser measurement system accuracy in linear measurement mode will be within: ± 0.5 ppm (k=2) - see the system manual for details.

Authorised signature	Signatory	Position	Date
	Chris Hunt	General Manager	29 April 2020

This certificate may not be reproduced other than in full, except with the prior written approval of:

Renishaw plc
 Laser & Calibration Products Division
 Bath Road, Woodchester
 Stroud
 Gloucestershire
 GL5 5EY
 United Kingdom
 Tel +44 (0) 1453 524524

Certificate number
8H1451-200429-00

L-9908-1128-02

Calibration notes

1. **Lasers (XM, XL, ML, HS and RLU)** are calibrated by comparison to a reference HeNe laser using an optical beat frequency technique. Reference lasers are routinely calibrated against an iodine-stabilised HeNe laser supplied by the National Physical Laboratory (NPL), or by a national standards laboratory. All frequency measurements are taken over a 1 hour period.
2. **Air pressure and relative humidity (RH) sensors are installed in a compensator (XC and RCU).**
The air pressure sensors are calibrated over 650 mbar to 1150 mbar range in a temperature controlled oven by direct comparison with a reference pressure meter. The RH sensors (where fitted) are certified by the manufacturer to be within specification. They are calibrated by comparison of the readings with those from a reference RH meter at a single applied humidity.
3. **Air and material temperature sensors (XC and RCU)** are calibrated by direct comparison with transfer platinum resistance thermometers (PRTs) in a temperature controlled water bath over 0 °C to 40 °C (50 °C for material sensor). The transfer PRTs are routinely calibrated against reference PRTs.
4. **Rotary axis calibrators (XR20)** are calibrated using a HeNe laser angular interferometer.
5. **Ballbar transducers (QC20-W and QC10)** are calibrated using a HeNe laser interferometer. The scale factor (QC10 only) is calculated and must be entered into the Renishaw application software prior to use.
6. **Ballbar calibrators** are calibrated by direct comparison with a reference ballbar calibrator (calibrated by a national standards laboratory) using a reference ballbar as a transfer standard. The measured values for the ballbar calibrator must be entered into the Renishaw application software prior to use.
7. **Traceability.** All the reference standards (listed overleaf) used in these calibrations are traceable either directly to major international metrology institutes who have signed the CIPM Mutual Recognition Agreement (e.g. NPL: UK; LNE: France; NIST: USA; PTB: Germany; NMIJ: Japan) or to a national accreditation body (e.g. UKAS: UK; A2LA: USA).
8. **Environment.** The equipment used for calibration is in a facility held between 15 °C and 25 °C.
9. **Uncertainty calculations.** The uncertainty calculations have been carried out according to the European Co-operation for Accreditation document EA-4/02.
10. **Quality accreditation.** All calibrations above are covered by Renishaw's ISO 9001 quality assurance system. The system is audited and certified by an accredited agency.
11. **Re-calibration.** Customers may wish to confirm that systems are performing within published specifications over time. If so, it is recommended that they should be periodically re-calibrated. Please note that compensators and temperature sensors are re-calibrated only at a single applied temperature, air pressure and humidity. Please refer to the appropriate system manual for further details.